Our Mission

Fostering a culture of innovation.
Leading healthcare solutions.
Simplifying complex care.

SOLEO HEALTH®
Simplifying Complex Care

SOLEO HEALTH®
Simplifying Complex Care

www.soleohealth.com

Administration of Medications Via The Curlin 6000 Pump

© 2025 Soleo Health. All Rights Reserved SPE1005-1.25

ADMINISTRATION OF MEDICATIONS VIA THE CURLIN 6000 PUMP

Your medication has been pre-mixed by the Soleo Health pharmacist and placed in a plastic bag. Store all medication as directed. Some medications need to be refrigerated for stability.

- Check the label of each bag for your name, name of the medication, volume and the expiration of the drug.
- Each dose should be free of cracks, leaks, floating objects and discoloration. If you have any doubts call your Soleo Health pharmacist.
- Remove one dose from your refrigerator one (1) hour prior to infusing to allow the medication to warm to room temperature. **NEVER HEAT OR MICROWAVE** the medication.
- Use each bag only once.
- Change the IV tubing as directed by your nurse.
- Tubing is used for a full 24 hours regardless of number of doses. If you have multiple doses due within a 24-hour period, you will be provided with caps to place on the end of the tubing between doses.

SCAN THE QR CODE BELOW TO VIEW THE VIDEO

STEP 1

Wash hands thoroughly using soap and water for at least 20 seconds. Dry with a paper towel. Use your towel to turn off the faucet.

STEP 2

Clean your workplace and gather your supplies.

1. MEDICATION BAG

2. PUMP

3. ALCOHOL WIPES

4. IV TUBING SET

5. FLUSH SYRINGES

6. SHARPS CONTAINER



STEP 3

Inspect your IV site for:

- REDNESS
- SWELLING
- TENDERNESS
- WARMTH

IF THERE IS A CHANGE, DO NOT INFUSE AND CONTACT YOUR NURSE

STEP 4

Inspect your medication.

- IS IT EXPIRED?
- IS YOUR NAME ON THE LABEL?
- IS IT THE RIGHT MEDICATION?
- ARE THERE ANY VISIBLE SPECKS OR FLOATING PARTICLES IN THE MEDICATION?

STEP 5

Remove tubing from wrapper and close clamps. Remove the protective covers from the IV bag administration port and the IV tubing spike. Push the spike in the administration port, making sure you don't touch the spike.

STEP 6

Break off yellow tab from the free flow valve.
Note: If your bags already have the tubing attached, the yellow tab may already be removed.





STEP 7

- Remove the air from the tubing by squeezing the yellow free flow clamp, allowing the tubing to fill with fluid.
- If your IV tubing has a filter, hold the filter upright until completely filled with fluid and you see fluid continue past the filter through the end of the line.

STEP 8

Open the door of the pump by lifting the lever on the top of the pump.

STEP 9

Swing the door arm of the pump to the open position.

STEP 10

Place the tubing into the pump by inserting the yellow clip into the slot marked with the yellow arrow.

STEP 11

Insert the blue clip into the slot marked with the blue arrow.

STEP 12

Close the door arm of the pump. Be careful not to pinch the tubing in the door lock.

STEP 13

Click the door lever down to lock the door in the closed position.

STEP 14

Swab the end of your needleless connector using an alcohol wipe. Allow to air dry. Do not blot or fan.



STEP 15

Remove the cap from the saline flush. Holding the syringe upright, gently pull the plunger back to loosen it, and then push the plunger up to remove the air from the syringe.

STEP 16

Attach saline flush syringe to catheter adapter by firmly pushing and twisting. Gently, but firmly flush catheter with ____mL of saline

Call your nurse if you are unable to flush the catheter or it is more difficult than usual.



STEP 17

Connect the IV tubing to the needleless connector and unclamp IV tubing.

STEP 18

Turn the pump on by pressing the orange ON/OFF key. The pump will perform a self-test and then beep three times.

STEP 19

A screen with **PROGRAM** will appear. The cursor (highlighted area) will be on **PROGRAM**.

STEP 20

Select **PROGRAM** by pressing the **YES/ENTER** key. The screen will change to display **REPEAT Rx** and **NEW PROGRAM**.

STEP 21

Your pump should be preprogrammed by your Soleo Health pharmacist or nurse. Move the cursor (highlighted area) to **REPEAT Rx** by pressing the down arrow key.

STEP 22

Select the **REPEAT Rx** option by pressing the **YES/ENTER** key.

STEP 23

Your programmed settings will appear in the next two display screens for you to review. The display screen will scroll automatically. The **RUN OPTIONS** screen will now appear. Press the **RUN/PAUSE** button to start your infusion.

STEP 24

Once the infusion is complete, push the **RUN/PAUSE** button, then push the **ON/OFF** button to turn off your pump.

STEP 25

Disconnect the tubing from your IV needleless connector.

STEP 26

Swab the IV needleless connector with alcohol and flush the IV catheter with saline flush solution as before. Swab the end of your IV needleless connector again with alcohol and allow to air dry. Follow with a heparin flush if instructed to do so using the prescribed amount.

Call your nurse if you are unable to flush your catheter or meet resistance.

STEP 27

To remove the tubing from the pump, lift the lever on the top of the pump and swing the door arm of the pump to the open position.

STEP 28

Remove the tubing from the pump.

STEP 29

Discard used medication bag and tubing in the appropriate waste container.

NOTES